

ENDO-MODEL[®]

4 DECADES OF CLINICAL
EXCELLENCE

PorEx Surface Modification:

Significant reduction of metal ions¹

Lower coefficient of friction
compared to cobalt chrome surfaces¹

ROTATION DURING FLEXION
STABLE IN EXTENSION

Smooth stop in full extension due to poly insert
Rotation dampened to 0 in full extension

¹ Internal technical report: Study of the influence of
TiNbN-coating on the ion release of CrCrMo-alloys in
SBF buffer simulator testing.

ENDO-MODEL®

4 DECADES OF CLINICAL
E X C E L L E N C E



B O N E C O N S E R V I N G
C O N E F R I E N D L Y D E S I G N

Minimal 14mm tibio-femoral joint resection.

Centrally located axis of rotation
offers reduced force across
patello-femoral joint

2° of hyper-extension featuring
anti-dislocation device

Tapered satin finished stems allow
for simple extraction